

常德比克曼生物科技有限公司
CHANGDE BKMAM BIOTECHNOLOGY CO.,LTD

产品质量控制报告
PRODUCT QUALITY REPROT


Product Name: Sterile HDPE Sampling Bottle 500mL

Lot No: 20240220

Expiry Date: 20270219

Item No.:110405072

1piece/bag, 80pcs/carton

Code	Test Item	Test Method/Standard	Test Result
1	Appearance	Black spots $\leq 0.2\text{mm}^2$, with ≤ 2 black spots on the same side.	Qualified
		No oil stains, no discoloration	Qualified
		No cracks/strains, no scratches	Qualified
		No color difference, no bubbles/perforations	Qualified
		No shrinkage/deformation, no shortage of materials	Qualified
		Gate/burr $\leq 0.3\text{mm}$	Qualified
2	Size	Same as drawing	Qualified
3	Print	printed content is correct and clear, without missing ink or ghosting	Qualified
		The printing is firm, and 3M tape is pasted three times without deinking	Qualified
4	Leakage	Inflate 10Kpa, leakage $\leq 10\text{pa}$	Qualified
5	Sealing	Evacuate -70kpa~-90kpa, no leakage for 3 minutes	Qualified
6	Package	The packaging and quantity are correct, and the packaging is intact	Qualified
7	Sterile	Ethylene oxide, sterilized for 6 hours, indicator card meets standard color, and sterile sampling bottle shape is stable	Qualified
8	RNase Testing	Rnase Pollution Detection Specification	Qualified
9	DNase Testing	DnasePollution Detection Specification	Qualified
Test Result			

操作人: 王桂华

复核: 周攀

技术主管: 熊志伟

日期: 20240220